


Data Path : W:\HPCHEM1\MSVOA X\Data\VX052918\  
 Data File : VX002084.D  
 Acq On : 29 May 2018 09:39  
 Operator : JC/MD  
 Sample : BFB  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BFB

Integration File: RTEINT.P

Method : W:\HPCHEM1\MSVOA\_X\METHOD\82X052418W.M  
 Title : SW846 8260  
 Last Update : Thu May 24 13:09:03 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1723

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 50          | 95           | 15           | 40           | 20.5      | 6168    | PASS             |
| 75          | 95           | 30           | 60           | 53.6      | 16121   | PASS             |
| 95          | 95           | 100          | 100          | 100.0     | 30088   | PASS             |
| 96          | 95           | 5            | 9            | 6.7       | 2017    | PASS             |
| 173         | 174          | 0.00         | 2            | 1.4       | 366     | PASS             |
| 174         | 95           | 50           | 100          | 87.7      | 26384   | PASS             |
| 175         | 174          | 5            | 9            | 7.9       | 2083    | PASS             |
| 176         | 174          | 95           | 101          | 98.0      | 25850   | PASS             |
| 177         | 176          | 5            | 9            | 6.3       | 1640    | PASS             |