

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053018\
 Data File : VX002135.D
 Acq On : 30 May 2018 12:26
 Operator : JC/MD
 Sample : VX0530WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0530WBL01

Quant Time: May 31 05:17:31 2018
 Quant Method : W:\HPCHEM1\MSVOA X\METHOD\624X052518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri May 25 02:17:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.02	128	52546	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.87	114	307125	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	268891	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.07	65	129097	31.22	ug/l	-0.01
Spiked Amount	30.000	Range	50 - 169	Recovery	=	104.07%
60) 4-Bromofluorobenzene	11.14	95	126546	27.49	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	91.63%
63) Toluene-d8	8.72	98	366592	30.59	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	101.97%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053018\
Data File : VX002135.D
Acq On : 30 May 2018 12:26
Operator : JC/MD
Sample : VX0530WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBL01

Quant Time: May 31 05:17:31 2018
Quant Method : W:\HPCHEM1\MSVOA X\METHOD\624X052518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri May 25 02:17:15 2018
Response via : Initial Calibration

