

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053018\
 Data File : VX002136.D
 Acq On : 30 May 2018 13:03
 Operator : JC/MD
 Sample : J3070-01 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 14B-1

Quant Time: May 31 05:20:37 2018
 Quant Method : W:\HPCHEM1\MSVOA X\METHOD\624X052518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri May 25 02:17:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.03	128	53143	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.87	114	314905	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	276184	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.07	65	133440	31.91	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	106.37%
60) 4-Bromofluorobenzene	11.14	95	130741	27.65	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	92.17%
63) Toluene-d8	8.72	98	378878	30.78	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	102.60%

Target Compounds					Ovalue
25) Chloroform	5.22	83	317768	54.49	ug/l 98
41) Bromodichloromethane	7.90	83	52611	10.51	ug/l 97
53) Dibromochloromethane	9.59	129	4445	1.05	ug/l 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053018\
Data File : VX002136.D
Acq On : 30 May 2018 13:03
Operator : JC/MD
Sample : J3070-01 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
14B-1

Quant Time: May 31 05:20:37 2018
Quant Method : W:\HPCHEM1\MSVOA X\METHOD\624X052518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri May 25 02:17:15 2018
Response via : Initial Calibration

