

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053018\
 Data File : VX002140.D
 Acq On : 30 May 2018 14:49
 Operator : JC/MD
 Sample : J3093-01 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 14B-1

Quant Time: May 31 05:29:17 2018
 Quant Method : W:\HPCHEM1\MSVOA X\METHOD\624X052518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri May 25 02:17:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.03	128	50197	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.87	114	284930	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	253024	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.07	65	126341	31.98	ug/l	0.00
Spiked Amount 30.000	Range 50 - 169		Recovery	=	106.60%	
60) 4-Bromofluorobenzene	11.14	95	116535	26.91	ug/l	0.00
Spiked Amount 30.000	Range 56 - 143		Recovery	=	89.70%	
63) Toluene-d8	8.72	98	341617	30.29	ug/l	0.00
Spiked Amount 30.000	Range 66 - 137		Recovery	=	100.97%	

Target Compounds

						Ovalue
25) Chloroform	5.22	83	307486	55.82	ug/l	99
41) Bromodichloromethane	7.90	83	48821	10.78	ug/l	99
53) Dibromochloromethane	9.59	129	4250	1.11	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053018\
Data File : VX002140.D
Acq On : 30 May 2018 14:49
Operator : JC/MD
Sample : J3093-01 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
14B-1

Quant Time: May 31 05:29:17 2018
Quant Method : W:\HPCHEM1\MSVOA X\METHOD\624X052518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri May 25 02:17:15 2018
Response via : Initial Calibration

