

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053018\
 Data File : VX002151.D
 Acq On : 30 May 2018 21:45
 Operator : JC/MD
 Sample : J3093-04 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 14B-4

Quant Time: May 31 05:52:43 2018
 Quant Method : W:\HPCHEM1\MSVOA X\METHOD\624X052518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri May 25 02:17:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.03	128	35429	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.87	114	190322	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	160663	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.07	65	86824	31.14	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	103.80%
60) 4-Bromofluorobenzene	11.14	95	72220	26.26	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	87.53%
63) Toluene-d8	8.72	98	227209	31.73	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	105.77%

Target Compounds

					Ovalue
25) Chloroform	5.22	83	253413	65.18	ug/l 97
41) Bromodichloromethane	7.90	83	39984	13.21	ug/l 100
53) Dibromochloromethane	9.59	129	3267	1.28	ug/l 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053018\
Data File : VX002151.D
Acq On : 30 May 2018 21:45
Operator : JC/MD
Sample : J3093-04 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
14B-4

Quant Time: May 31 05:52:43 2018
Quant Method : W:\HPCHEM1\MSVOA X\METHOD\624X052518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri May 25 02:17:15 2018
Response via : Initial Calibration

