

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053018\
 Data File : VX002152.D
 Acq On : 30 May 2018 22:11
 Operator : JC/MD
 Sample : J3192-01 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 14B-1

Quant Time: May 31 05:53:44 2018
 Quant Method : W:\HPCHEM1\MSVOA X\METHOD\624X052518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri May 25 02:17:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.03	128	49251	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.87	114	284188	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	247406	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.07	65	123439	31.85	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	106.17%
60) 4-Bromofluorobenzene	11.14	95	114535	27.04	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	90.13%
63) Toluene-d8	8.72	98	337378	30.60	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	102.00%


Target Compounds					Ovalue
25) Chloroform	5.22	83	283784	52.51 ug/l	98
41) Bromodichloromethane	7.90	83	44212	9.78 ug/l	97
53) Dibromochloromethane	9.59	129	3567	0.93 ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053018\
Data File : VX002152.D
Acq On : 30 May 2018 22:11
Operator : JC/MD
Sample : J3192-01 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
14B-1

Quant Time: May 31 05:53:44 2018
Quant Method : W:\HPCHEM1\MSVOA X\METHOD\624X052518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri May 25 02:17:15 2018
Response via : Initial Calibration





#1

Bromochloromethane

Concen: 30.00 ug/l

RT: 5.03 min Scan# 729

Delta R.T. 0.01 min

Lab File: VX002152.D

Acq: 30 May 2018 22:11

Instrument :

MSVOA_X

ClientSampleId :

14B-1


Tot Ion:128 Resp: 49251

Ion Ratio Lower Upper

128 100

49 174.4 0.0 423.0




130 129.9 0.0 324.3



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 49.00 (48.70 to 49.70): VX0

Ion 130.00 (129.70 to 130.70): V



#25

Chloroform

Concen: 52.51 ug/l

RT: 5.22 min Scan# 760

Delta R.T. -0.01 min

Lab File: VX002152.D


Acq: 30 May 2018 22:11

Tgt Ion: 83 Resp: 283784

Ion Ratio Lower Upper



83 100


85 66.4 51.9 77.9



Abundance Ion 83.00 (82.70 to 83.70): VX0

Ion 85.00 (84.70 to 85.70): VX0









#27
 1,2-Dichloroethane-d4
 Concen: 31.85 ug/l
 RT: 6.07 min Scan# 900
 Delta R.T. -0.01 min
 Lab File: VX002152.D
 Acq: 30 May 2018 22:11




Instrument :
 MSVOA_X
 ClientSampled :
 14B-1


Tot Ion: 65 Resp: 123439
 Ion Ratio Lower Upper
 65 100
 67 51.8 40.9 61.3



#28
 1,4-Difluorobenzene
 Concen: 30.00 ug/l
 RT: 6.87 min Scan# 1030
 Delta R.T. -0.00 min
 Lab File: VX002152.D
 Acq: 30 May 2018 22:11

Tgt Ion: 114 Resp: 284188
 Ion Ratio Lower Upper
 114 100
 63 21.4 17.1 25.7
 88 15.7 12.6 19.0





#41

Bromodichloromethane

Concen: 9.78 ug/l

RT: 7.90 min Scan# 1200

Delta R.T. -0.01 min

Lab File: VX002152.D

Acq: 30 May 2018 22:11

Instrument :

MSVOA_X

ClientSampleId :

14B-1


Tot Ion: 83 Resp: 44212

Ion Ratio Lower Upper

83 100

85 65.2 49.8 74.8




127 9.2 7.3 10.9



Abundance Ion 83.00 (82.70 to 83.70): VX002152.D (-1193) (-)

Ion 85.00 (84.70 to 85.70): VX002152.D (-1193) (-)

Ion 127.00 (126.70 to 127.70): VX002152.D (-1193) (-)



#53

Dibromochloromethane

Concen: 0.93 ug/l

RT: 9.59 min Scan# 1476

Delta R.T. -0.00 min

Lab File: VX002152.D


Acq: 30 May 2018 22:11

Tgt Ion: 129 Resp: 3567

Ion Ratio Lower Upper



129 100


127 76.3 38.5 115.3



Abundance Ion 129.00 (128.70 to 129.70): VX002152.D (-1394) (-)

Ion 127.00 (126.70 to 127.70): VX002152.D (-1394) (-)









#57
Chlorobenzene-d5
Concen: 30.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. -0.01 min
Lab File: VX002152.D
Acq: 30 May 2018 22:11




Instrument :
MSVOA_X
ClientSampleId :
14B-1


Tot Ion	Ratio	Lower	Upper
117	100		
82	55.5	44.3	66.5
119	32.2	25.7	38.5



#60
4-Bromofluorobenzene
Concen: 27.04 ug/l
RT: 11.14 min Scan# 1731
Delta R.T. -0.00 min
Lab File: VX002152.D
Acq: 30 May 2018 22:11

Tgt Ion	Ratio	Lower	Upper
95	100		
174	90.7	71.8	107.6
176	88.3	70.1	105.1





#63

Toluene-d8

Concen: 30.60 ug/l

RT: 8.72 min Scan# 1334

Delta R.T. -0.00 min

Lab File: VX002152.D

Acq: 30 May 2018 22:11

Instrument :

MSVOA_X

ClientSampleId :

14B-1

Tot Ion: 98 Resp: 337378

Ion Ratio Lower Upper

98 100

100 63.6 50.8 76.2

