

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053118\
 Data File : VX002174.D
 Acq On : 31 May 2018 13:38
 Operator : JC/MD
 Sample : J3173-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 W-1

Quant Time: May 31 16:30:08 2018
 Quant Method : W:\HPCHEM1\MSVOA X\METHOD\624X052518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri May 25 02:17:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.03	128	48074	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.87	114	256351	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	208584	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.07	65	116122	30.70	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	102.33%
60) 4-Bromofluorobenzene	11.14	95	96488	27.02	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	90.07%
63) Toluene-d8	8.72	98	290652	31.26	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	104.20%

Target Compounds

					Ovalue
15) Acetone	2.45	58	10122	23.805 ug/l	91
34) Benzene	6.15	78	15934	1.471 ug/l	98
62) Toluene	8.79	91	23314	2.189 ug/l	99
68) o-Xylene	10.70	106	4930	1.083 ug/l	96
91) Naphthalene	13.83	128	14217	1.062 ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053118\
Data File : VX002174.D
Acq On : 31 May 2018 13:38
Operator : JC/MD
Sample : J3173-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-1

Quant Time: May 31 16:30:08 2018
Quant Method : W:\HPCHEM1\MSVOA X\METHOD\624X052518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri May 25 02:17:15 2018
Response via : Initial Calibration

