

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053118\
 Data File : VX002175.D
 Acq On : 31 May 2018 14:05
 Operator : JC/MD
 Sample : J3192-03 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 14B-3

Quant Time: May 31 16:16:03 2018
 Quant Method : W:\HPCHEM1\MSVOA X\METHOD\624X052518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri May 25 02:17:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.03	128	52326	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.87	114	304788	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	270544	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.07	65	129201	31.38	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	104.60%
60) 4-Bromofluorobenzene	11.14	95	130386	28.15	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	93.83%
63) Toluene-d8	8.72	98	364354	30.22	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	100.73%

Target Compounds

					Ovalue
15) Acetone	2.45	58	1678	3.626	ug/l 81
25) Chloroform	5.22	83	264691	46.098	ug/l 99
41) Bromodichloromethane	7.90	83	42656	8.802	ug/l 98
53) Dibromochloromethane	9.59	129	3393	0.828	ug/l 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053118\
Data File : VX002175.D
Acq On : 31 May 2018 14:05
Operator : JC/MD
Sample : J3192-03 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
14B-3

Quant Time: May 31 16:16:03 2018
Quant Method : W:\HPCHEM1\MSVOA X\METHOD\624X052518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri May 25 02:17:15 2018
Response via : Initial Calibration

