

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053118\
 Data File : VX002179.D
 Acq On : 31 May 2018 16:07
 Operator : JC/MD
 Sample : VX0531WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0531WBL02

Quant Time: May 31 17:12:09 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	250151	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	351851	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	315716	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	170407	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	239479	53.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.12%	
35) Dibromofluoromethane	5.51	113	179903	51.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.26%	
50) Toluene-d8	8.72	98	706183	55.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.88%	
62) 4-Bromofluorobenzene	11.14	95	248083	49.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.56%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053118\
Data File : VX002179.D
Acq On : 31 May 2018 16:07
Operator : JC/MD
Sample : VX0531WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0531WBL02

Quant Time: May 31 17:12:09 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

