

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053118\
 Data File : VX002188.D
 Acq On : 31 May 2018 20:50
 Operator : JC/MD
 Sample : J3118-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW718-180522

Quant Time: Jun 01 04:07:44 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	264440	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	382879	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	348010	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	191385	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	244051	51.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.26%	
35) Dibromofluoromethane	5.51	113	178573	46.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.28%	
50) Toluene-d8	8.72	98	706705	50.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.96%	
62) 4-Bromofluorobenzene	11.14	95	250007	46.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.20%	
Target Compounds						
12) 1,1-Dichloroethene	2.38	96	1039	0.337	ug/l	88
27) cis-1,2-Dichloroethene	4.60	96	16221	4.036	ug/l	88
44) Trichloroethene	7.22	130	45195	9.940	ug/l	96
55) 1,1,2-Trichloroethane	9.22	97	1180	0.284	ug/l #	86
64) Tetrachloroethene	9.34	164	8264	1.656	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053118\
Data File : VX002188.D
Acq On : 31 May 2018 20:50
Operator : JC/MD
Sample : J3118-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW718-180522

Quant Time: Jun 01 04:07:44 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

