

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053118\
 Data File : VX002191.D
 Acq On : 31 May 2018 22:10
 Operator : JC/MD
 Sample : I.BLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 I.BLK

Quant Time: Jun 01 03:56:19 2018
 Quant Method : W:\HPCHEM1\MSVOA X\METHOD\624X052518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri May 25 02:17:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.03	128	51829	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.87	114	295794	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	253978	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.07	65	126303	30.97	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	103.23%
60) 4-Bromofluorobenzene	11.14	95	117489	27.02	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	90.07%
63) Toluene-d8	8.72	98	349432	30.87	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	102.90%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053118\
Data File : VX002191.D
Acq On : 31 May 2018 22:10
Operator : JC/MD
Sample : I.BLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
I.BLK

Quant Time: Jun 01 03:56:19 2018
Quant Method : W:\HPCHEM1\MSVOA X\METHOD\624X052518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri May 25 02:17:15 2018
Response via : Initial Calibration

