

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053118\
 Data File : VX002205.D
 Acq On : 01 Jun 2018 08:15
 Operator : JC/MD
 Sample : J3119-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW715-180522

Quant Time: Jun 01 09:10:20 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	257175	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	367447	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	332249	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	185379	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	218813	47.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.20%	
35) Dibromofluoromethane	5.51	113	160102	43.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.14%	
50) Toluene-d8	8.72	98	640129	48.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.24%	
62) 4-Bromofluorobenzene	11.14	95	226073	43.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.88%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	5585	1.429	ug/l	39
44) Trichloroethene	7.22	130	18984	4.351	ug/l	96
64) Tetrachloroethene	9.34	164	4100	0.861	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053118\
Data File : VX002205.D
Acq On : 01 Jun 2018 08:15
Operator : JC/MD
Sample : J3119-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW715-180522

Quant Time: Jun 01 09:10:20 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

