


Data Path : W:\HPCHEM1\MSVOA X\DATA\VX053118\
Data File : VX002197.D
Acq On : 01 Jun 2018 01:15
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Title : SW846 8260
Last Update : Thu May 24 13:09:03 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1723

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.3	2210	PASS
75	95	30	60	53.4	5546	PASS
95	95	100	100	100.0	10380	PASS
96	95	5	9	7.2	750	PASS
173	174	0.00	2	0.9	89	PASS
174	95	50	100	91.1	9451	PASS
175	174	5	9	7.6	717	PASS
176	174	95	101	96.3	9100	PASS
177	176	5	9	6.9	627	PASS