

Data Path : W:\HPCHEM1\MSVOA_X\DATA\VX060118\
 Data File : VX002219.D
 Acq On : 01 Jun 2018 18:36
 Operator : JC/MD
 Sample : J3118-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW733-180521

Quant Time: Jun 02 03:51:20 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	222743	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	319447	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	283306	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	151517	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	219815	55.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.42%	
35) Dibromofluoromethane	5.51	113	161638	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
50) Toluene-d8	8.72	98	633802	54.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.60%	
62) 4-Bromofluorobenzene	11.14	95	215554	47.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.28%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	15992	4.724	ug/l	95
44) Trichloroethene	7.22	130	42333	11.160	ug/l	92
55) 1,1,2-Trichloroethane	9.22	97	1286	0.372	ug/l	96
64) Tetrachloroethene	9.34	164	7492	1.844	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA_X\DATA\VX060118\
Data File : VX002219.D
Acq On : 01 Jun 2018 18:36
Operator : JC/MD
Sample : J3118-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW733-180521

Quant Time: Jun 02 03:51:20 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

