

Data Path : W:\HPCHEM1\MSVOA_X\DATA\VX060118\
 Data File : VX002227.D
 Acq On : 01 Jun 2018 22:10
 Operator : JC/MD
 Sample : J3151-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW353-180523

Quant Time: Jun 02 04:27:25 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	227421	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	338047	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	306540	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	165379	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	227035	55.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.70%	
35) Dibromofluoromethane	5.51	113	165936	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
50) Toluene-d8	8.72	98	656243	53.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.24%	
62) 4-Bromofluorobenzene	11.14	95	227058	47.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.84%	
Target Compounds						
19) Methyl tert-butyl Ether	3.21	73	11160	1.119	ug/l	93
27) cis-1,2-Dichloroethene	4.60	96	4987	1.443	ug/l	41
44) Trichloroethene	7.22	130	3762	0.937	ug/l	88
64) Tetrachloroethene	9.34	164	7634	1.737	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA_X\DATA\VX060118\
Data File : VX002227.D
Acq On : 01 Jun 2018 22:10
Operator : JC/MD
Sample : J3151-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW353-180523

Quant Time: Jun 02 04:27:25 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

