

Data Path : W:\HPCHEM1\MSVOA_X\DATA\VX060118\
 Data File : VX002229.D
 Acq On : 01 Jun 2018 23:03
 Operator : JC/MD
 Sample : J3151-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW183-180523

Quant Time: Jun 02 04:35:09 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	230323	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	337172	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	306056	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	166584	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	257176	62.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	124.94%	
35) Dibromofluoromethane	5.51	113	186586	55.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.68%	
50) Toluene-d8	8.72	98	730171	59.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.64%	
62) 4-Bromofluorobenzene	11.14	95	249929	52.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.66%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	24570	7.019	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA_X\DATA\VX060118\
Data File : VX002229.D
Acq On : 01 Jun 2018 23:03
Operator : JC/MD
Sample : J3151-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW183-180523

Quant Time: Jun 02 04:35:09 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

