

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX060218\  
 Data File : VX002239.D  
 Acq On : 02 Jun 2018 15:10  
 Operator : JC/MD  
 Sample : J3151-08  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 SS052MW183-180523

Quant Time: Jun 04 13:03:11 2018  
 Quant Method : W:\HPCHEM1\MSVOA\_X\METHOD\82X052418W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu May 24 13:09:03 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units   | Dev(Min)  |
|-----------------------------|-------|------|----------|-------|---------|-----------|
| 1) Pentafluorobenzene       | 5.67  | 168  | 261027   | 50.00 | ug/l    | 0.00      |
| 34) 1,4-Difluorobenzene     | 6.87  | 114  | 367649   | 50.00 | ug/l    | 0.00      |
| 63) Chlorobenzene-d5        | 10.12 | 117  | 334241   | 50.00 | ug/l    | 0.00      |
| 72) 1,4-Dichlorobenzene-d4  | 12.09 | 152  | 190306   | 50.00 | ug/l    | 0.00      |
| System Monitoring Compounds |       |      |          |       |         |           |
| 33) 1,2-Dichloroethane-d4   | 6.07  | 65   | 233117   | 49.96 | ug/l    | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 99.92%  |           |
| 35) Dibromofluoromethane    | 5.51  | 113  | 171337   | 46.60 | ug/l    | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 93.20%  |           |
| 50) Toluene-d8              | 8.72  | 98   | 669343   | 50.29 | ug/l    | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 100.58% |           |
| 62) 4-Bromofluorobenzene    | 11.14 | 95   | 236126   | 45.34 | ug/l    | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 90.68%  |           |
| Target Compounds            |       |      |          |       |         |           |
| 27) cis-1,2-Dichloroethene  | 4.60  | 96   | 25072    | 6.320 | ug/l    | Qvalue 95 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX060218\  
Data File : VX002239.D  
Acq On : 02 Jun 2018 15:10  
Operator : JC/MD  
Sample : J3151-08  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
SS052MW183-180523

Quant Time: Jun 04 13:03:11 2018  
Quant Method : W:\HPCHEM1\MSVOA\_X\METHOD\82X052418W.M  
Quant Title : SW846 8260  
QLast Update : Thu May 24 13:09:03 2018  
Response via : Initial Calibration

