

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX060218\
 Data File : VX002240.D
 Acq On : 02 Jun 2018 15:37
 Operator : JC/MD
 Sample : J3151-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW180-180523

Quant Time: Jun 04 05:54:56 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	237837	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	352558	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	312604	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	165171	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	227216	53.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.90%	
35) Dibromofluoromethane	5.51	113	166995	47.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.74%	
50) Toluene-d8	8.72	98	659646	51.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.36%	
62) 4-Bromofluorobenzene	11.14	95	225673	45.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.38%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	14644	4.051	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX060218\
Data File : VX002240.D
Acq On : 02 Jun 2018 15:37
Operator : JC/MD
Sample : J3151-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW180-180523

Quant Time: Jun 04 05:54:56 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

