

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX060218\
 Data File : VX002243.D
 Acq On : 02 Jun 2018 16:57
 Operator : JC/MD
 Sample : J3119-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW737-180521

Quant Time: Jun 04 06:10:07 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 24 13:09:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	258305	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	371802	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	346739	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	191085	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	228843	49.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.12%	
35) Dibromofluoromethane	5.51	113	164758	44.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.62%	
50) Toluene-d8	8.72	98	657316	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
62) 4-Bromofluorobenzene	11.14	95	228406	43.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.74%	
Target Compounds						
12) 1,1-Dichloroethene	2.37	96	3976	1.321	ug/l	94
44) Trichloroethene	7.21	130	985	0.223	ug/l	85
64) Tetrachloroethene	9.34	164	2601	0.523	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX060218\
Data File : VX002243.D
Acq On : 02 Jun 2018 16:57
Operator : JC/MD
Sample : J3119-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW737-180521

Quant Time: Jun 04 06:10:07 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X052418W.M
Quant Title : SW846 8260
QLast Update : Thu May 24 13:09:03 2018
Response via : Initial Calibration

