

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
 Data File : VX041695.D
 Acq On : 03 Jun 2024 11:33
 Operator : JC/MD
 Sample : VX0603WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBL01

Quant Time: Jun 04 00:44:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143582	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	219058	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217545	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	111449	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	82241	46.702	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.400%
35) Dibromofluoromethane	5.385	113	69968	46.197	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.400%
50) Toluene-d8	8.647	98	263060	49.709	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.420%
62) 4-Bromofluorobenzene	11.079	95	99258	50.818	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
Data File : VX041695.D
Acq On : 03 Jun 2024 11:33
Operator : JC/MD
Sample : VX0603WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBL01

Quant Time: Jun 04 00:44:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

