

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
 Data File : VX041702.D
 Acq On : 03 Jun 2024 14:20
 Operator : JC/MD
 Sample : P2660-02 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-08

Quant Time: Jun 04 00:47:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143635	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	220791	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218419	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110829	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83059	47.149	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.300%		
35) Dibromofluoromethane	5.385	113	69255	45.367	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.740%		
50) Toluene-d8	8.647	98	260225	48.787	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.580%		
62) 4-Bromofluorobenzene	11.079	95	99672	50.630	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	101.260%		
Target Compounds						
						Qvalue
16) Acetone	2.392	43	2498	0.542	ug/l	95
20) Methylene Chloride	2.782	84	1064	0.675	ug/l	86
40) Benzene	6.038	78	7413	1.283	ug/l	98
52) Toluene	8.720	92	7375	2.011	ug/l	96
67) Ethyl Benzene	10.195	91	186671	24.988	ug/l	97
68) m/p-Xylenes	10.299	106	259010	87.154	ug/l	94
69) o-Xylene	10.640	106	130206	44.041	ug/l	95
73) Isopropylbenzene	10.964	105	10393	1.512	ug/l	100
78) n-propylbenzene	11.305	91	26218	3.473	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	52782	9.196	ug/l	97
84) 1,2,4-Trimethylbenzene	11.750	105	217496	37.399	ug/l	92
95) Naphthalene	13.774	128	76297	10.627	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
Data File : VX041702.D
Acq On : 03 Jun 2024 14:20
Operator : JC/MD
Sample : P2660-02 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-08

Quant Time: Jun 04 00:47:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

