

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
 Data File : VX041708.D
 Acq On : 03 Jun 2024 16:41
 Operator : JC/MD
 Sample : P2660-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11

Quant Time: Jun 04 00:50:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	145986	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	221909	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216201	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	106757	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81021	45.251	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.500%
35) Dibromofluoromethane	5.391	113	69378	45.219	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.440%
50) Toluene-d8	8.646	98	261850	48.844	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.680%
62) 4-Bromofluorobenzene	11.079	95	97408	49.231	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.460%
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	6728	5.869	ug/l	97
30) Chloroform	5.098	83	11192	4.309	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
Data File : VX041708.D
Acq On : 03 Jun 2024 16:41
Operator : JC/MD
Sample : P2660-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11

Quant Time: Jun 04 00:50:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

