

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
 Data File : VX041712.D
 Acq On : 03 Jun 2024 18:14
 Operator : JC/MD
 Sample : P2687-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 72-11977

Quant Time: Jun 04 00:52:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	136853	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	207171	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101855	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	78832	46.967	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.940%
35) Dibromofluoromethane	5.385	113	65182	45.506	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.020%
50) Toluene-d8	8.647	98	248277	49.607	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.220%
62) 4-Bromofluorobenzene	11.079	95	94037	50.908	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.820%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	86418	116.000	ug/l	97
18) Methyl Acetate	2.715	43	3350	2.090	ug/l	92
20) Methylene Chloride	2.782	84	3663	2.439	ug/l	99
25) 2-Butanone	4.581	43	20228	19.391	ug/l	96
43) Isopropyl Acetate	6.348	43	110798	33.528	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
Data File : VX041712.D
Acq On : 03 Jun 2024 18:14
Operator : JC/MD
Sample : P2687-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
72-11977

Quant Time: Jun 04 00:52:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

