

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
 Data File : VX041713.D
 Acq On : 03 Jun 2024 18:38
 Operator : JC/MD
 Sample : P2687-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 72-11998

Quant Time: Jun 04 00:53:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	138541	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	208976	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	210435	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102157	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78196	46.020	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.040%
35) Dibromofluoromethane	5.385	113	65272	45.176	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.360%
50) Toluene-d8	8.647	98	251830	49.882	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.760%
62) 4-Bromofluorobenzene	11.079	95	95172	51.077	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.160%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	87096	115.466	ug/l	97
18) Methyl Acetate	2.709	43	4607	2.840	ug/l	100
20) Methylene Chloride	2.782	84	3923	2.580	ug/l	93
25) 2-Butanone	4.581	43	20086	19.020	ug/l	96
43) Isopropyl Acetate	6.349	43	111043	33.312	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
Data File : VX041713.D
Acq On : 03 Jun 2024 18:38
Operator : JC/MD
Sample : P2687-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
72-11998

Quant Time: Jun 04 00:53:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

