

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
 Data File : VX041715.D
 Acq On : 03 Jun 2024 19:25
 Operator : JC/MD
 Sample : P2657-18 100X
 Misc : 4.99g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 5-AB-VOC

Quant Time: Jun 04 00:54:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	134770	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	204440	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	199271	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101704	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	76878	46.511	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	93.020%		
35) Dibromofluoromethane	5.385	113	63263	44.757	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	89.520%		
50) Toluene-d8	8.647	98	237569	48.102	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.200%		
62) 4-Bromofluorobenzene	11.079	95	90992	49.917	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	99.840%		
Target Compounds						
18) Methyl Acetate	2.709	43	2118	1.342	ug/l	94
95) Naphthalene	13.780	128	11890	1.805	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
Data File : VX041715.D
Acq On : 03 Jun 2024 19:25
Operator : JC/MD
Sample : P2657-18 100X
Misc : 4.99g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
5-AB-VOC

Quant Time: Jun 04 00:54:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

