

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
 Data File : VX041717.D
 Acq On : 03 Jun 2024 20:12
 Operator : JC/MD
 Sample : P2657-10 100X
 Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 3-AB-VOC

Quant Time: Jun 04 00:55:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	138891	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	208563	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202269	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103527	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78581	46.130	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.260%
35) Dibromofluoromethane	5.385	113	65078	45.131	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.260%
50) Toluene-d8	8.646	98	244392	48.505	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.000%
62) 4-Bromofluorobenzene	11.079	95	92294	49.631	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.260%
Target Compounds						
95) Naphthalene	13.774	128	155369	23.167	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
Data File : VX041717.D
Acq On : 03 Jun 2024 20:12
Operator : JC/MD
Sample : P2657-10 100X
Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
3-AB-VOC

Quant Time: Jun 04 00:55:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

