

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
 Data File : VX041718.D
 Acq On : 03 Jun 2024 20:35
 Operator : JC/MD
 Sample : P2657-06 100X
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 2-AB-VOC

Quant Time: Jun 04 00:55:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	134080	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	203374	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200188	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103395	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77587	47.181	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.360%
35) Dibromofluoromethane	5.391	113	63319	45.031	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.060%
50) Toluene-d8	8.647	98	236672	48.171	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.340%
62) 4-Bromofluorobenzene	11.079	95	92311	50.906	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.820%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2863	1.277	ug/l	95
68) m/p-Xylenes	10.305	106	3730	1.369	ug/l	100
69) o-Xylene	10.646	106	3072	1.134	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	8578	1.581	ug/l	91
95) Naphthalene	13.774	128	464885	69.408	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
Data File : VX041718.D
Acq On : 03 Jun 2024 20:35
Operator : JC/MD
Sample : P2657-06 100X
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
2-AB-VOC

Quant Time: Jun 04 00:55:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

