

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
 Data File : VX041719.D
 Acq On : 03 Jun 2024 20:59
 Operator : JC/MD
 Sample : P2657-02 100X
 Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1-AB-VOC

Quant Time: Jun 04 00:55:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	131196	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	200463	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	197165	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101419	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75102	46.674	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.340%
35) Dibromofluoromethane	5.385	113	61493	44.368	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.740%
50) Toluene-d8	8.647	98	233445	48.204	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.400%
62) 4-Bromofluorobenzene	11.079	95	90100	50.409	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.820%
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	2828	1.314	ug/l	97
95) Naphthalene	13.774	128	352568	53.665	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
Data File : VX041719.D
Acq On : 03 Jun 2024 20:59
Operator : JC/MD
Sample : P2657-02 100X
Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1-AB-VOC

Quant Time: Jun 04 00:55:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

