



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060324\
Data File : VX041692.D
Acq On : 03 Jun 2024 09:42
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Title : SW846 8260
Last Update : Thu May 30 07:36:24 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.4	8432	PASS
75	95	30	60	50.6	21960	PASS
95	95	100	100	100.0	43384	PASS
96	95	5	9	7.0	3036	PASS
173	174	0.00	2	1.3	541	PASS
174	95	50	100	93.7	40656	PASS
175	174	5	9	7.6	3107	PASS
176	174	95	101	97.2	39536	PASS
177	176	5	9	6.9	2735	PASS