

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060418\
 Data File : VX002272.D
 Acq On : 04 Jun 2018 20:34
 Operator : JC/MD
 Sample : J3234-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF6-0060

Quant Time: Jun 05 06:02:47 2018
 Quant Method : Y:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 05 03:22:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	269843	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	397231	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	376088	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	221202	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	191536	49.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.22%	
35) Dibromofluoromethane	5.51	113	139228	44.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.14%	
50) Toluene-d8	8.72	98	573779	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
62) 4-Bromofluorobenzene	11.14	95	212281	45.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.96%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060418\
Data File : VX002272.D
Acq On : 04 Jun 2018 20:34
Operator : JC/MD
Sample : J3234-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF6-0060

Quant Time: Jun 05 06:02:47 2018
Quant Method : Y:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060418W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 05 03:22:35 2018
Response via : Initial Calibration

