

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022413.D
 Acq On : 04 Jun 2021 15:35
 Operator : JC/MD
 Sample : M2563-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 33213-14

Quant Time: Jun 04 15:56:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	65787	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	138200	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122788	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45567	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	57007	56.175	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.360%
35) Dibromofluoromethane	5.391	113	41659	47.534	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.060%
50) Toluene-d8	8.653	98	169788	50.293	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.580%
62) 4-Bromofluorobenzene	11.085	95	57580	46.144	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.280%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	13786	31.199	ug/l	94
20) Methylene Chloride	2.794	84	6145	6.263	ug/l #	85
43) Isopropyl Acetate	6.342	43	13949	5.444	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022413.D
Acq On : 04 Jun 2021 15:35
Operator : JC/MD
Sample : M2563-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
33213-14

Quant Time: Jun 04 15:56:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

