

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022416.D
 Acq On : 04 Jun 2021 16:40
 Operator : JC/MD
 Sample : M2512-09 4X
 Misc : 1.20g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-4-6

Quant Time: Jun 05 03:20:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	61612	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	125241	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	114166	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53411	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50728	53.375	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 106.740%			
35) Dibromofluoromethane	5.397	113	37101	46.714	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 93.420%			
50) Toluene-d8	8.653	98	155240	50.742	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.480%			
62) 4-Bromofluorobenzene	11.085	95	61145	54.071	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 108.140%			
Target Compounds						Qvalue
17) Carbon Disulfide	2.501	76	12300	6.651	ug/l #	95
31) Cyclohexane	5.483	56	41341	34.615	ug/l	89
39) Methylcyclohexane	7.385	83	81339	61.928	ug/l	98
40) Benzene	6.044	78	51815	14.541	ug/l	97
52) Toluene	8.726	92	76423	33.867	ug/l	99
67) Ethyl Benzene	10.201	91	44604	10.202	ug/l	98
68) m/p-Xylenes	10.305	106	107567	65.101	ug/l	99
69) o-Xylene	10.646	106	38207	23.775	ug/l	98
73) Isopropylbenzene	10.963	105	9804	2.169	ug/l	95
78) n-propylbenzene	11.305	91	22759	4.298	ug/l	96
80) 1,3,5-Trimethylbenzene	11.457	105	72089	19.241	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	145978	38.708	ug/l	99
85) sec-Butylbenzene	11.896	105	9067	1.986	ug/l	89
86) p-Isopropyltoluene	12.012	119	11330	2.938	ug/l	92
95) Naphthalene	13.780	128	79494	19.030	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022416.D
Acq On : 04 Jun 2021 16:40
Operator : JC/MD
Sample : M2512-09 4X
Misc : 1.20g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-4-6

Quant Time: Jun 05 03:20:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

