

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060421\  
 Data File : VX022418.D  
 Acq On : 04 Jun 2021 17:24  
 Operator : JC/MD  
 Sample : M2583-01ME  
 Misc : 5.05g/10mL/100uL/5.0mL/MSVOA\_X/MEOH  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 39-41SME

Quant Time: Jun 05 03:21:01 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X051921W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed May 19 14:23:10 2021  
 Response via : Initial Calibration

| Compound                    | R.T. QIon |       | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|-----------|-------|----------|----------|------------|----------|
| -----                       |           |       |          |          |            |          |
| Internal Standards          |           |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.550     | 168   | 63093    | 50.000   | ug/l       | #-0.01   |
| 34) 1,4-Difluorobenzene     | 6.763     | 114   | 129485   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.061    | 117   | 114309   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024    | 152   | 50382    | 50.000   | ug/l       | 0.00     |
|                             |           |       |          |          |            |          |
| System Monitoring Compounds |           |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.964     | 65    | 50654    | 52.046   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 78 - 117 | Recovery | = 104.100% |          |
| 35) Dibromofluoromethane    | 5.397     | 113   | 37492    | 45.659   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 75 - 124 | Recovery | = 91.320%  |          |
| 50) Toluene-d8              | 8.653     | 98    | 159426   | 50.402   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 92 - 112 | Recovery | = 100.800% |          |
| 62) 4-Bromofluorobenzene    | 11.085    | 95    | 57489    | 49.172   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 83 - 123 | Recovery | = 98.340%  |          |
|                             |           |       |          |          |            |          |
| Target Compounds            |           |       |          |          |            | Qvalue   |
| 16) Acetone                 | 2.434     | 43    | 4898     | 11.558   | ug/l       | 95       |
| 52) Toluene                 | 8.720     | 92    | 10003    | 4.288    | ug/l       | 96       |
| 68) m/p-Xylenes             | 10.305    | 106   | 5007     | 3.027    | ug/l       | 94       |
| 69) o-Xylene                | 10.646    | 106   | 2112     | 1.313    | ug/l       | 71       |
| 84) 1,2,4-Trimethylbenzene  | 11.756    | 105   | 5613     | 1.578    | ug/l       | 97       |
| -----                       |           |       |          |          |            |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060421\  
Data File : VX022418.D  
Acq On : 04 Jun 2021 17:24  
Operator : JC/MD  
Sample : M2583-01ME  
Misc : 5.05g/10mL/100uL/5.0mL/MSVOA\_X/MEOH  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
39-41SME

Quant Time: Jun 05 03:21:01 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X051921W.M  
Quant Title : SW846 8260  
QLast Update : Wed May 19 14:23:10 2021  
Response via : Initial Calibration

