

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022424.D
 Acq On : 04 Jun 2021 19:46
 Operator : JC/MD
 Sample : M2580-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 18-B1(2-4)

Quant Time: Jun 05 03:22:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	66071	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	138530	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	123497	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49351	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	55256	54.216	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.440%
35) Dibromofluoromethane	5.397	113	41294	47.006	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.020%
50) Toluene-d8	8.653	98	170420	50.360	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.720%
62) 4-Bromofluorobenzene	11.085	95	60519	48.384	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.760%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	12987	29.264	ug/l	99
20) Methylene Chloride	2.794	84	3823	3.880	ug/l #	86
43) Isopropyl Acetate	6.355	43	11760	4.579	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022424.D
Acq On : 04 Jun 2021 19:46
Operator : JC/MD
Sample : M2580-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
18-B1(2-4)

Quant Time: Jun 05 03:22:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

