

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022430.D
 Acq On : 04 Jun 2021 22:30
 Operator : JC/MD
 Sample : VX0604WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0604WBL02

Quant Time: Jun 05 03:24:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	65902	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	138598	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119886	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47256	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54549	53.659	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.320%
35) Dibromofluoromethane	5.397	113	41683	47.425	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.860%
50) Toluene-d8	8.653	98	168202	49.680	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.360%
62) 4-Bromofluorobenzene	11.085	95	58036	46.376	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022430.D
Acq On : 04 Jun 2021 22:30
Operator : JC/MD
Sample : VX0604WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0604WBL02

Quant Time: Jun 05 03:24:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

