

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022433.D
 Acq On : 04 Jun 2021 23:41
 Operator : JC/MD
 Sample : PB136806TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136806TB

Quant Time: Jun 05 03:26:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	59087	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	122675	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	106973	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39354	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	48649	53.375	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.740%
35) Dibromofluoromethane	5.397	113	36074	46.371	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.740%
50) Toluene-d8	8.653	98	149044	49.735	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.480%
62) 4-Bromofluorobenzene	11.085	95	51116	46.148	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.300%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	12683	31.957	ug/l	98
18) Methyl Acetate	2.715	43	1506	1.427	ug/l #	79
20) Methylene Chloride	2.794	84	3146	3.570	ug/l #	85
43) Isopropyl Acetate	6.355	43	12265	5.392	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022433.D
Acq On : 04 Jun 2021 23:41
Operator : JC/MD
Sample : PB136806TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136806TB

Quant Time: Jun 05 03:26:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

