

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022439.D
 Acq On : 05 Jun 2021 02:02
 Operator : JC/MD
 Sample : PB136806ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136806ZHE#06

Quant Time: Jun 05 06:21:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	62193	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	130563	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116731	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44818	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	52685	54.916	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 109.840%			
35) Dibromofluoromethane	5.391	113	38334	46.299	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 92.600%			
50) Toluene-d8	8.653	98	161587	50.663	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.320%			
62) 4-Bromofluorobenzene	11.085	95	55316	46.923	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 93.840%			
Target Compounds					Qvalue	
16) Acetone	2.392	43	6495	15.548	ug/l	100
18) Methyl Acetate	2.715	43	1522	1.370	ug/l #	77
20) Methylene Chloride	2.800	84	4680	5.046	ug/l #	95
43) Isopropyl Acetate	6.355	43	14582	6.024	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022439.D
Acq On : 05 Jun 2021 02:02
Operator : JC/MD
Sample : PB136806ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136806ZHE#06

Quant Time: Jun 05 06:21:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

