

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022440.D
 Acq On : 05 Jun 2021 02:25
 Operator : JC/MD
 Sample : PB136806ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136806ZHE#07

Quant Time: Jun 05 06:21:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	65781	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	136524	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	123851	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47270	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53776	52.996	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.000%
35) Dibromofluoromethane	5.397	113	40360	46.617	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.240%
50) Toluene-d8	8.653	98	168964	50.663	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.320%
62) 4-Bromofluorobenzene	11.085	95	58383	47.362	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.720%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	6220	14.078	ug/l	# 83
18) Methyl Acetate	2.715	43	2122	1.806	ug/l	# 84
20) Methylene Chloride	2.794	84	4114	4.194	ug/l	90
43) Isopropyl Acetate	6.348	43	14436	5.703	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022440.D
Acq On : 05 Jun 2021 02:25
Operator : JC/MD
Sample : PB136806ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136806ZHE#07

Quant Time: Jun 05 06:21:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

