

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022441.D
 Acq On : 05 Jun 2021 02:49
 Operator : JC/MD
 Sample : PB136806ZHE#08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136806ZHE#08

Quant Time: Jun 05 06:22:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	64452	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	136682	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	125984	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49315	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	55039	55.359	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 110.720%			
35) Dibromofluoromethane	5.397	113	40842	47.120	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 94.240%			
50) Toluene-d8	8.653	98	170752	51.140	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 102.280%			
62) 4-Bromofluorobenzene	11.085	95	60332	48.887	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.780%			
Target Compounds					Qvalue	
16) Acetone	2.392	43	6154	14.215	ug/l	94
18) Methyl Acetate	2.709	43	1851	1.608	ug/l #	86
20) Methylene Chloride	2.794	84	4318	4.492	ug/l #	93
43) Isopropyl Acetate	6.348	43	14256	5.625	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022441.D
Acq On : 05 Jun 2021 02:49
Operator : JC/MD
Sample : PB136806ZHE#08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136806ZHE#08

Quant Time: Jun 05 06:22:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

