

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022442.D
 Acq On : 05 Jun 2021 03:13
 Operator : JC/MD
 Sample : PB136806ZHE#09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136806ZHE#09

Quant Time: Jun 05 06:22:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	65889	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	137922	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	123187	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49040	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	55258	54.367	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 108.740%			
35) Dibromofluoromethane	5.391	113	40926	46.792	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 93.580%			
50) Toluene-d8	8.653	98	169642	50.351	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.700%			
62) 4-Bromofluorobenzene	11.085	95	60793	48.817	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.640%			
Target Compounds						
16) Acetone	2.386	43	7109	16.063	ug/l	99
18) Methyl Acetate	2.715	43	1748	1.485	ug/l #	82
20) Methylene Chloride	2.794	84	4678	4.761	ug/l	93
43) Isopropyl Acetate	6.348	43	14322	5.601	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022442.D
Acq On : 05 Jun 2021 03:13
Operator : JC/MD
Sample : PB136806ZHE#09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136806ZHE#09

Quant Time: Jun 05 06:22:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

