

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022443.D
 Acq On : 05 Jun 2021 03:36
 Operator : JC/MD
 Sample : PB136806ZHE#10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136806ZHE#10

Quant Time: Jun 05 06:22:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	64073	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	135043	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124407	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49521	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54844	55.489	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 110.980%			
35) Dibromofluoromethane	5.397	113	40577	47.382	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 94.760%			
50) Toluene-d8	8.653	98	171424	51.964	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 103.920%			
62) 4-Bromofluorobenzene	11.085	95	59612	48.889	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.780%			
Target Compounds						Qvalue
16) Acetone	2.386	43	6602	15.340	ug/l	91
18) Methyl Acetate	2.709	43	1955	1.708	ug/l #	80
20) Methylene Chloride	2.794	84	4049	4.237	ug/l	94
43) Isopropyl Acetate	6.348	43	14074	5.621	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022443.D
Acq On : 05 Jun 2021 03:36
Operator : JC/MD
Sample : PB136806ZHE#10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136806ZHE#10

Quant Time: Jun 05 06:22:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

