

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022445.D
 Acq On : 05 Jun 2021 04:23
 Operator : JC/MD
 Sample : PB136806ZHE#12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136806ZHE#12

Quant Time: Jun 05 06:23:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	65261	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	134263	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122406	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46889	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54909	54.544	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.080%
35) Dibromofluoromethane	5.397	113	39730	46.663	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.320%
50) Toluene-d8	8.653	98	169336	51.630	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.260%
62) 4-Bromofluorobenzene	11.085	95	58778	48.485	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.980%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	6991	15.949	ug/l	96
18) Methyl Acetate	2.709	43	1946	1.669	ug/l #	84
20) Methylene Chloride	2.794	84	4663	4.791	ug/l	99
43) Isopropyl Acetate	6.348	43	14409	5.788	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022445.D
Acq On : 05 Jun 2021 04:23
Operator : JC/MD
Sample : PB136806ZHE#12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136806ZHE#12

Quant Time: Jun 05 06:23:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

