

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022446.D
 Acq On : 05 Jun 2021 04:47
 Operator : JC/MD
 Sample : PB136806ZHE#13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136806ZHE#13

Quant Time: Jun 05 06:23:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	64097	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	134966	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	123507	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49227	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	55610	56.243	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.480%
35) Dibromofluoromethane	5.397	113	40205	46.974	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.940%
50) Toluene-d8	8.653	98	169553	51.427	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.860%
62) 4-Bromofluorobenzene	11.085	95	60328	49.505	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.000%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	6873	15.964	ug/l	# 85
18) Methyl Acetate	2.715	43	2006	1.752	ug/l	95
20) Methylene Chloride	2.794	84	4041	4.227	ug/l	98
43) Isopropyl Acetate	6.355	43	14548	5.814	ug/l	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022446.D
Acq On : 05 Jun 2021 04:47
Operator : JC/MD
Sample : PB136806ZHE#13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136806ZHE#13

Quant Time: Jun 05 06:23:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

