

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022450.D
 Acq On : 05 Jun 2021 06:21
 Operator : JC/MD
 Sample : PB136806ZHE#17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136806ZHE#17

Quant Time: Jun 05 07:46:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	63293	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	133919	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	119439	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44546	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54137	55.449	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 110.900%			
35) Dibromofluoromethane	5.391	113	39855	46.930	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 93.860%			
50) Toluene-d8	8.653	98	165821	50.688	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.380%			
62) 4-Bromofluorobenzene	11.085	95	55996	46.309	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 92.620%			
Target Compounds						
16) Acetone	2.386	43	12012	28.255	ug/l	99
18) Methyl Acetate	2.715	43	1719	1.520	ug/l #	85
20) Methylene Chloride	2.794	84	3315	3.512	ug/l #	82
43) Isopropyl Acetate	6.348	43	11866	4.779	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022450.D
Acq On : 05 Jun 2021 06:21
Operator : JC/MD
Sample : PB136806ZHE#17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136806ZHE#17

Quant Time: Jun 05 07:46:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

