

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022451.D
 Acq On : 05 Jun 2021 06:44
 Operator : JC/MD
 Sample : PB136806ZHE#18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136806ZHE#18

Quant Time: Jun 05 07:47:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	64766	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	136104	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	117100	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45314	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53233	53.283	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 106.560%			
35) Dibromofluoromethane	5.397	113	39517	45.785	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 91.560%			
50) Toluene-d8	8.653	98	165683	49.833	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 99.660%			
62) 4-Bromofluorobenzene	11.085	95	56441	45.928	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 91.860%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	12162	27.957	ug/l	98
18) Methyl Acetate	2.715	43	1909	1.650	ug/l #	71
20) Methylene Chloride	2.800	84	3418	3.539	ug/l #	86
43) Isopropyl Acetate	6.348	43	11913	4.721	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022451.D
Acq On : 05 Jun 2021 06:44
Operator : JC/MD
Sample : PB136806ZHE#18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136806ZHE#18

Quant Time: Jun 05 07:47:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

