

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022461.D
 Acq On : 05 Jun 2021 11:02
 Operator : JC/MD
 Sample : M2580-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 18-B2(2-4)

Manual Integrations
 APPROVED

MMDadoda
 6/7/2021 6:51:16 PM

Quant Time: Jun 07 03:21:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	64273	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	134030	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	120452	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	48041	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54735	55.207	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 110.420%			
35) Dibromofluoromethane	5.397	113	39885	46.926	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 93.860%			
50) Toluene-d8	8.653	98	165152	50.442	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.880%			
62) 4-Bromofluorobenzene	11.085	95	57753	47.723	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 95.440%			
Target Compounds						
16) Acetone	2.385	43	14174	32.832	ug/l	Qvalue 92
20) Methylene Chloride	2.794	84	3171	3.308	ug/l	# 75
25) 2-Butanone	4.592	43	3501m	5.013	ug/l	
43) Isopropyl Acetate	6.348	43	10940	4.402	ug/l	# 90
95) Naphthalene	13.780	128	9488	2.525	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022461.D
Acq On : 05 Jun 2021 11:02
Operator : JC/MD
Sample : M2580-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 57 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
18-B2(2-4)

Manual Integrations
APPROVED

MMDadoda
6/7/2021 6:51:16 PM

Quant Time: Jun 07 03:21:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

