

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022462.D
 Acq On : 05 Jun 2021 11:26
 Operator : JC/MD
 Sample : M2580-23
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 18-B2(4-6)

Quant Time: Jun 07 03:21:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	63216	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	128838	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	114078	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	40842	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53245	54.602	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.200%	
35) Dibromofluoromethane	5.397	113	38482	47.100	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 94.200%	
50) Toluene-d8	8.653	98	158333	50.308	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.620%	
62) 4-Bromofluorobenzene	11.085	95	53673	46.139	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.280%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	14598	34.380	ug/l	96
20) Methylene Chloride	2.794	84	3395	3.601	ug/l #	94
43) Isopropyl Acetate	6.349	43	10515	4.402	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022462.D
Acq On : 05 Jun 2021 11:26
Operator : JC/MD
Sample : M2580-23
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 58 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
18-B2(4-6)

Quant Time: Jun 07 03:21:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

