

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022468.D
 Acq On : 05 Jun 2021 13:47
 Operator : JC/MD
 Sample : M2580-29
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 64 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 18-B4(4-6)

Quant Time: Jun 07 03:23:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	61114	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	129817	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	114562	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46711	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52495	55.684	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.360%
35) Dibromofluoromethane	5.397	113	34127	41.455	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	82.900%
50) Toluene-d8	8.653	98	157783	49.755	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.500%
62) 4-Bromofluorobenzene	11.085	95	56789	48.449	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.900%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	38553	93.920	ug/l	93
20) Methylene Chloride	2.794	84	3651	4.006	ug/l #	87
25) 2-Butanone	4.568	43	4342	6.538	ug/l #	85
95) Naphthalene	13.780	128	10818	2.961	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022468.D
Acq On : 05 Jun 2021 13:47
Operator : JC/MD
Sample : M2580-29
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 64 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
18-B4(4-6)

Quant Time: Jun 07 03:23:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

