

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022474.D
 Acq On : 05 Jun 2021 16:08
 Operator : JC/MD
 Sample : M2589-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 70 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 18-B7(0-2)

Quant Time: Jun 07 03:25:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	61842	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	128387	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116659	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47409	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51171	53.641	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 107.280%			
35) Dibromofluoromethane	5.397	113	38106	46.804	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 93.600%			
50) Toluene-d8	8.653	98	158121	50.417	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.840%			
62) 4-Bromofluorobenzene	11.085	95	57618	49.704	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 99.400%			
Target Compounds						
16) Acetone	2.386	43	11810	28.432	ug/l	88
20) Methylene Chloride	2.794	84	4688	5.083	ug/l	91
25) 2-Butanone	4.568	43	4077	6.067	ug/l #	74
43) Isopropyl Acetate	6.348	43	12342	5.185	ug/l #	92
68) m/p-Xylenes	10.299	106	6086	3.605	ug/l	96
69) o-Xylene	10.646	106	10335	6.294	ug/l	93
73) Isopropylbenzene	10.963	105	7634	1.903	ug/l	95
78) n-propylbenzene	11.305	91	10636	2.263	ug/l	98
80) 1,3,5-Trimethylbenzene	11.457	105	46069	13.853	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	125736	37.562	ug/l	99
85) sec-Butylbenzene	11.896	105	9380	2.314	ug/l	85
86) p-Isopropyltoluene	12.012	119	7565	2.210	ug/l	94
95) Naphthalene	13.780	128	130323	35.148	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022474.D
Acq On : 05 Jun 2021 16:08
Operator : JC/MD
Sample : M2589-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 70 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
18-B7(0-2)

Quant Time: Jun 07 03:25:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

