

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022454.D
 Acq On : 05 Jun 2021 07:55
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 05 08:15:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	79	0.00
2 T	Dichlorodifluoromethane	0.608	0.581	4.4	72	0.00
3 P	Chloromethane	0.656	0.687	-4.7	80	0.00
4 C	Vinyl Chloride	0.694	0.715	-3.0#	78	0.00
5 T	Bromomethane	0.404	0.410	-1.5	83	-0.01
6 T	Chloroethane	0.413	0.441	-6.8	79	0.00
7 T	Trichlorofluoromethane	0.998	1.013	-1.5	77	0.00
8 T	Diethyl Ether	0.429	0.412	4.0	76	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.592	0.583	1.5	77	0.00
10 T	Methyl Iodide	0.637	0.685	-7.5	80	0.00
11 T	Tert butyl alcohol	0.176	0.204	-15.9	93	-0.01
12 CM	1,1-Dichloroethene	0.573	0.615	-7.3#	83	0.00
13 T	Acrolein	0.117	0.112	4.3	80	0.00
14 T	Allyl chloride	1.053	1.103	-4.7	80	0.00
15 T	Acrylonitrile	0.347	0.415	-19.6	91	0.00
16 T	Acetone	0.336	0.387	-15.2	94	0.00
17 T	Carbon Disulfide	1.501	1.457	2.9	79	0.00
18 T	Methyl Acetate	0.893	1.065	-19.3	95	0.00
19 T	Methyl tert-butyl Ether	1.963	2.149	-9.5	84	0.00
20 T	Methylene Chloride	0.746	0.759	-1.7	85	0.00
21 T	trans-1,2-Dichloroethene	0.623	0.679	-9.0	82	0.00
22 T	Diisopropyl ether	2.027	2.330	-14.9	89	0.00
23 T	Vinyl Acetate	1.952	2.253	-15.4	88	0.00
24 P	1,1-Dichloroethane	1.190	1.327	-11.5	86	0.00
25 T	2-Butanone	0.543	0.662	-21.9	94	0.00
26 T	2,2-Dichloropropane	1.056	0.562	46.8#	41#	0.00
27 T	cis-1,2-Dichloroethene	0.758	0.817	-7.8	81	0.00
28 T	Bromochloromethane	0.577	0.612	-6.1	85	0.00
29 T	Tetrahydrofuran	0.323	0.385	-19.2	93	0.00
30 C	Chloroform	1.229	1.317	-7.2#	83	0.00
31 T	Cyclohexane	0.969	1.042	-7.5	84	0.00
32 T	1,1,1-Trichloroethane	1.050	1.095	-4.3	80	0.00
33 S	1,2-Dichloroethane-d4	0.771	0.813	-5.4	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.317	0.306	3.5	86	0.00
36 T	1,1-Dichloropropene	0.462	0.471	-1.9	83	0.00
37 T	Ethyl Acetate	0.567	0.609	-7.4	91	0.00
38 T	Carbon Tetrachloride	0.452	0.429	5.1	74	0.00
39 T	Methylcyclohexane	0.524	0.494	5.7	79	0.00
40 TM	Benzene	1.423	1.484	-4.3	85	0.00
41 T	Methacrylonitrile	0.303	0.327	-7.9	89	0.00
42 TM	1,2-Dichloroethane	0.498	0.529	-6.2	87	0.00
43 T	Isopropyl Acetate	0.927	1.012	-9.2	90	0.00
44 TM	Trichloroethene	0.354	0.340	4.0	78	0.00
45 C	1,2-Dichloropropane	0.363	0.386	-6.3#	83	0.00
46 T	Dibromomethane	0.246	0.255	-3.7	82	0.00
47 T	Bromodichloromethane	0.493	0.483	2.0	77	0.00
48 T	Methyl methacrylate	0.427	0.474	-11.0	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
 Data File : VX022454.D
 Acq On : 05 Jun 2021 07:55
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 05 08:15:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.010	-25.0	93	-0.02
50 S	Toluene-d8	1.221	1.152	5.7	85	0.00
51 T	4-Methyl-2-Pentanone	0.555	0.639	-15.1	92	0.00
52 CM	Toluene	0.901	0.901	0.0	81	0.00
53 T	t-1,3-Dichloropropene	0.569	0.503	11.6	71	0.00
54 T	cis-1,3-Dichloropropene	0.611	0.552	9.7	73	0.00
55 T	1,1,2-Trichloroethane	0.366	0.364	0.5	80	0.00
56 T	Ethyl methacrylate	0.591	0.630	-6.6	84	0.00
57 T	1,3-Dichloropropane	0.624	0.666	-6.7	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.320	0.308	3.8	80	0.00
59 T	2-Hexanone	0.418	0.484	-15.8	92	0.00
60 T	Dibromochloromethane	0.356	0.328	7.9	72	0.00
61 T	1,2-Dibromoethane	0.364	0.385	-5.8	83	0.00
62 S	4-Bromofluorobenzene	0.451	0.438	2.9	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.328	0.299	8.8	73	0.00
65 PM	Chlorobenzene	1.041	1.034	0.7	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.362	0.353	2.5	76	0.00
67 C	Ethyl Benzene	1.915	1.927	-0.6#	81	0.00
68 T	m/p-Xylenes	0.724	0.710	1.9	80	0.00
69 T	o-Xylene	0.704	0.702	0.3	81	0.00
70 T	Styrene	1.153	1.173	-1.7	80	0.00
71 P	Bromoform	0.250	0.220	12.0	66	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
73 T	Isopropylbenzene	4.232	4.247	-0.4	79	0.00
74 T	N-amyl acetate	2.009	2.361	-17.5	90	0.00
75 P	1,1,2,2-Tetrachloroethane	1.403	1.547	-10.3	86	0.00
76 T	1,2,3-Trichloropropane	1.234	1.368	-10.9	84	0.00
77 T	Bromobenzene	0.897	0.902	-0.6	78	0.00
78 T	n-propylbenzene	4.957	5.075	-2.4	79	0.00
79 T	2-Chlorotoluene	3.006	2.997	0.3	79	0.00
80 T	1,3,5-Trimethylbenzene	3.507	3.576	-2.0	78	0.00
81 T	trans-1,4-Dichloro-2-butene	0.521	0.401	23.0	58	0.00
82 T	4-Chlorotoluene	3.464	3.464	0.0	80	0.00
83 T	tert-Butylbenzene	3.364	3.362	0.1	77	0.00
84 T	1,2,4-Trimethylbenzene	3.530	3.611	-2.3	78	0.00
85 T	sec-Butylbenzene	4.275	4.277	-0.0	76	0.00
86 T	p-Isopropyltoluene	3.610	3.515	2.6	75	0.00
87 T	1,3-Dichlorobenzene	1.730	1.719	0.6	77	0.00
88 T	1,4-Dichlorobenzene	1.788	1.715	4.1	77	0.00
89 T	n-Butylbenzene	3.340	3.212	3.8	73	0.00
90 T	Hexachloroethane	0.597	0.537	10.1	67	0.00
91 T	1,2-Dichlorobenzene	1.662	1.716	-3.2	80	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.300	0.314	-4.7	81	0.00
93 T	1,2,4-Trichlorobenzene	1.003	0.978	2.5	76	0.00
94 T	Hexachlorobutadiene	0.391	0.317	18.9	63	0.00
95 T	Naphthalene	3.910	4.174	-6.8	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060421\
Data File : VX022454.D
Acq On : 05 Jun 2021 07:55
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 05 08:15:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.011	0.983	2.8	76	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 5